

Conditions of Acceptability - When installed in the end-product, consideration shall be given to the following:

1. This component has been judged on the basis of the required spacings in the Standard for Safety of Information Technology Equipment, Including Electrical Business Equipment, CSA C22.2 No. 60950, Third Edition, dated December 1, 2000 and the Standard for Medical Electrical Equipment, UL2601-1/CAN/CSA C22.2 No. 601.1, Second Edition **as well as for the Standard for Power Units other than Class 2 UL1012, sixth edition.**

2. The products were tested on a 16 A branch circuit. If used on a branch circuit greater than this, additional testing may be necessary.

3. All secondary output circuits are SELV according to the Standard for Safety of Information Technology Equipment, Including Electrical Business Equipment, CSA C22.2 No. 60950, Third Edition, dated December 1, 2000 and are not hazardous energy levels.

4. The terminals and connectors are suitable for field wiring.

5. The power supply shall be properly bonded to the main protective earthing termination in the end product.

6. Bonding terminals provided on this equipment have not been evaluated as protective earthing terminals. (Tab located on enclosure of Model PSR 230/24-1.3 and B 0310110)

7. Magnetic device(s) (e.g. transformer, inductor) L3 employ(s) an (OBJY2), electrical insulation system designated **Class F or alternate an electrical insulation system designated Class B.**